

Micro Centrifuge

Max RCF : 21,000 xg or Better

Max Speed : 14,000 rpm or Better

Max Noise Level : 56 dBA (ventilated); 50 dBA (refrigerated)

Acceleration/Deceleration Rates : 11/12 seconds (ventilated); 10/12 seconds (refrigerated)

Time Set Range : 1 min - 99 min; 1 min increments + HOLD mode

Certifications : CSA certified, CE marked, IVD compliant, Certified biocontainment with Biocontainment certification by a Third Party

Technical Standards : IEC 61010-1, IEC 61010-2, IEC 61010-020

Should be supplied with:

24 x 1.5/2.0 mL rotor

with ClickSeal biocontainment lid certified by a 3rd Party

10 x 5 mL rotor

with ClickSeal biocontainment lid certified by a 3rd Party for Nucleic Acid Centrifugation

Personal Centrifuge to be included